

Rollator Breaks need adjusting

If the wheels won't lock after the patient squeezes the breaks, simply tighten them. To do so, loosen/unscrew the bolt that connects the break wire to the handle (Lefty loosey/Righty tighty). The more this bolt is loosened, the tighter the breaks will become.

Also, if the breaks are too tight, you can tighten the bolt, which will loosen the breaks.

